

## Certificate of Analysis

# Reference Material 2202-SA Secondary Alumina

| Analyte                        | Certified value    | Uncertainty |
|--------------------------------|--------------------|-------------|
|                                | Mass fraction in % |             |
| Na₂O                           | 0,78               | ± 0,26      |
| SiO <sub>2</sub>               | 0,018              | ± 0,008     |
| Fe <sub>2</sub> O <sub>3</sub> | 0,036              | ± 0,005     |
| ZnO                            | 0,0031             | ± 0,0012    |
| CaO                            | 0,040              | ± 0,005     |
| TiO <sub>2</sub>               | 0,0029             | ± 0,0007    |
| $P_2O_5$                       | < 0,005            |             |

| Analyte                                  | Certified value    | Uncertainty |
|--|--------------------|-------------|
|  | Mass fraction in % |             |
| MnO                                      | < 0,002            |             |
| $V_2O_5$                                 | < 0,005            |             |
| $Ga_2O_3$                                | 0,012              | ± 0,004     |
| С  | 0,22               | ± 0,08      |
| S  | < 0,5              |             |
| F  | 2,6                | ± 1,0       |
| $\alpha$ -Al <sub>2</sub> O <sub>3</sub> | 4,8                | ± 1,4       |

Date of issue: August 31, 2022





### Reference material 2202-SA

#### Description

The base material for this secondary Alumina reference material has been taken directly from the production cycle after gas treatment process of an Aluminum smelter. This is process material which can be regarded as sufficiently homogeneous.

Grainsize D90: 130  $\pm$  19  $\mu$ m

This reference material was certified in an interlaboratory test of 7 laboratories. The values given in this certificate are taken from the evaluation of the interlaboratory test.

The uncertainties were estimated at a 95 % confidence level, showing both the contribution of homogeneity and the uncertainties of the analytical methods used. The uncertainty values were calculated from the reproducibility standard deviations of the ILT with a coverage factor k = 2.

The certified values are given as oxides on an equivalent weight basis and assume stoichiometry in the oxide form listed.

#### Recommended use

This reference material is intended for the verification of analytical methods such as XRF, XRD, ICP OES or calibration of analytical instruments.

#### Storage information

This reference material should be stored in a dry and clean environment at room temperature, e.g. storage in desiccator.

#### Methods used for characterization

XRF, XRD and combustion analysis.

#### Disclaimer

We inspect and test to the best of our knowledge and belief and assume no further liability for the accuracy of the inspection and test.

#### **Contact**

For more information see www.metallogie.de/ringversuche/

META<L>LOGIE, c/o pdv-software GmbH Im Schleeke 50, 38642 Goslar, Germany

E-Mail: ringversuch@metallogie.de